


MicroBooNE Blue Epoxy Test 4/17/2013 – 4/19/2013

1. Zero test.
2. Warm test.



Pen Name	Description	Value	Eng Units	High Over Range	Low Over Range
— Luke.PRM_LIFETIME.F_CV	Luke.PRM_LIFETIME.F_CV	0.00211	sec	0.01403	0.00171
— Luke.PRM_IMPURITIES.F_CV	Luke.PRM_IMPURITIES.F_CV	0.0712	imps	0.0878	0.0107
— PAB_S1_LT_13_AR_REAL_F_CV	Luke Argon Level Probe	23.1	inches	24.9	20.7
— PAB_S1_LT_10_N2_F_CV	Luke Condenser LN2 Level Probe (F_CV)	4.6	inches	14.5	1.0
— PAB_S1_PT_69_V_REAL_F_CV	Seal Monitor Vacuum	0.0010	Torr	0.1162	0.0001
— PAB_S1_AE_600_AR_REAL_F_CV	Luke Halo (F_CV)	20.8	ppb	38.0	4.3

4/12/2013 7:17:45 AM 4/19/2013 11:22:45 PM