


MicroBooNE Silver Galvanized Wire Test 8/6/2013 – 8/14/2013

1. Zero test (room temp.)
2. Liquid test
3. Room Temp. Test



Pen Name	Description	Value	Eng Units	High Over Range	Low Over Range
— Luke.PRM_LIFETIME.F_CV	Luke.PRM_LIFETIME.F_CV	0.00354	sec	0.02000	0.00187
— Luke.PRM_IMPURITIES.F_CV	Luke.PRM_IMPURITIES.F_CV	0.0424	Imps	0.0804	0.0075
— PAB_S1.LT_13.AR.REAL.F_CV	Luke Argon Level Probe	18.0	inches	21.3	17.6
— PAB_S1.LT_10.N2.F_CV	Luke Condenser LN2 Level Probe (F_CV)	7.5	inches	14.5	0.7
— PAB_S1.CRYOSTAT_PRESSURE...	Luke Vapor Pressure	12.7	psig	17.6	11.5
— PAB_S1.TE_411_CS.REAL.C_F_CV	TC on Luke top flange (F_CV)	15.6	C	19.9	6.6
— PAB_S1.PT_69_V.REAL.F_CV	Seal Monitor Vacuum	0.0016	Torr	6.2227	0.0004
— PAB_S1.AE_600.AR.REAL.F_CV	Luke Halo (F_CV)	28.4	ppb	483.8	3.0

Water – black pen

Lifetime – green pen

Liquid level – yellow pen